

www.QuantumMicrowave.com

QMC-CRYOATT-20BLK


Thermalized Non-Magnetic Crystalline Quartz Cryogenic Attenuators


- Frequency: DC to 18 GHz
- Attenuation: 20 dB
- Minimal change in attenuation with temperature
- Excellent RF Match
- Non-superconducting microstrip pad
- NiCr Crystalline Quartz conductive dielectric
- Silver Epoxy bonded
- Machined OFHC block, gold plated
- SMA Male and SMA Female Bulkhead
- Stackable mechanical design
- 100% Non-Magnetic




| Parameter | Specification | Notes |
|-------------------------------|--------------------|----------|
| Frequency Range | DC to 18 GHz | |
| Attenuation Setting @5GHz | 20dB +/- 1.0 dB | |
| Insertion Loss variation with | 20dB: 0.5dB typ. | See plot |
| frequency | | |
| Return Loss, 20 dB atten. | -20dB typ. | See plot |
| Connectors | SMA (m-bulkhead f) | |
| Max Power Handling | 100 mW | |

Measured Data (typical)





Outline Drawing [inches]



Quantum Microwave Components Phone 857-499-0071, Fax 888-893-7188

Magnetism Test information:

- Before the measurement, the devices are put in proximity (below 1cm) to a magnet for about 5 seconds, to induce magnetization of eventual magnetic material.
- The maghetic fied is measured over the surface of the device and along different axis. The distance between probe and device is below 1cm. The values reported in the picture is the maximum detected.
- All the measurements are in Gauss [Gs]
- The measurements have been done with a TD8620 magnetometer. Sensitivity is 0.1 Gs.
- The values are reported on the arrow corresponding to the direction of the maghetic fied

Part Number: QMC-CRYOATT-20NM

Description: 20 dB Cryogenic Non-Magnetic Attenuator **Comments:** The Field of the device is below the 0.1 Gs

